

Preparatory **PLUMBING** specifications and guidelines



HYDROTONE GEISHA PETITE TUB



- The following information is to assist the plumbing contractor in preparation for the delivery and installation of the Hydrotone Geisha Petite Tub.
- All plumbing work performed **must** be in accordance with local and national codes.
- An installation and user guide is supplied with the equipment.

GENERAL DESCRIPTION

- The Hydrotone Geisha Petite Tub is a stand alone hydrotherapy tub.
- Fitted surround panels enclose the tub. One removable panel allows access to the plumbing connections under the tub.
- A 6" (150mm) clearance between the tub chassis and the floor is allowed in the tub's design. The clearance allows for correct routing of the plumbing.

	REQUIREMENTS
Hot and Cold Water Supply – ¾" (19mm)	<ul style="list-style-type: none"> • The ¾" (19mm) hot and cold lines should exit the floor as shown in the floor plan provided - install shut-off valves to lines. • Hot and cold water supply lines are to be connected to the ¾" (19mm) female unions located directly under the water fill spout - underside of the tub • Ensure water lines are purged of sand and grit etc... prior to final connection to the unit. It is recommended that ¾" (19mm) true-flow in-line serviceable water strainers be fitted. • NOTE: ½" (13mm) water supply lines and 1 ½" (38mm) drain may be used – however, this will result in slower filling and draining of the tub.
Water Drainage – 2" (50mm)	<ul style="list-style-type: none"> • A 2" (50mm) drain should be located as shown in the floor plan provided. • The tubs drain is to be connected to this floor drain. • An additional floor drain is required in the room to cater for drainage of the floor in the event that flooding occurs.
Water Consumption	<ul style="list-style-type: none"> • The Hydrotone Geisha Petite Tub has a working capacity of approximately 45 gallons (170 litres) • Assuming a 50/50 mix of hot and cold water, this equates to a Hot water requirement of approximately 23 gallons (85 litres) for each bathing session. • A bathing session may last 40 minutes, therefore based upon the number of expected bathing sessions daily; an appropriate hot water supply needs to be installed.

Manufacturer reserves the right to amend specifications without prior notice

Updated: 20 October 2004